
Midmark M11 - v1.0.5
Total Cycles: 114

Name: RITTER2
Sterilizer ID: V2313228
06/17/2021 01:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.8 F 0.0
2:00 153.9 F 2.0
4:00 160.6 F 4.8
6:00 182.2 F 8.8
8:00 202.6 F 13.1
10:00 214.7 F 6.5
12:00 226.9 F 10.9
14:00 229.8 F 7.7
16:00 241.8 F 13.5
18:00 243.4 F 12.6
20:00 255.2 F 18.7
22:00 265.5 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 272.4 F 29.7
2:00 272.4 F 30.1
3:00 272.0 F 29.4
4:00 272.6 F 30.1

Min 271.7 F 29.3
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.3 F 0.0
5:00 239.6 F 0.0
10:00 227.8 F 0.0
15:00 231.1 F 0.0
20:00 224.9 F 0.0
25:00 228.7 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:49
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

06/17/2021 02:50 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____