

-----  
Midmark M11 - v1.0.5  
Total Cycles: 857

Name: RITTER2  
Sterilizer ID: V2313228  
03/02/2022 11:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.4 F 0.0  
2:00 187.2 F 3.5  
4:00 191.8 F 8.1  
6:00 209.9 F 13.6  
8:00 221.6 F 8.4  
10:00 233.2 F 11.5  
12:00 239.1 F 11.6  
14:00 239.1 F 10.5  
16:00 252.5 F 17.4  
18:00 264.4 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.6 F 29.9  
2:00 271.3 F 29.2  
3:00 272.3 F 30.1  
4:00 272.2 F 29.6  
  
Min 271.2 F 29.2  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 218.9 F 0.0  
10:00 233.1 F 0.0  
15:00 226.4 F 0.0  
20:00 223.1 F 0.0  
25:00 221.2 F 0.0  
30:00 221.2 F 0.0  
Drying Complete

FILLING: 1:26  
HEATING: 19:56  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:58:28  
CYCLE COMPLETE

03/02/2022 12:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_