
Midmark M11 - v1.0.5
Total Cycles: 3916

Name: RITTER2
Sterilizer ID: V2313228
09/09/2024 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.2 F 0.0
2:00 108.0 F 0.0
4:00 113.4 F 0.3
6:00 156.6 F 2.5
8:00 175.6 F 5.2
10:00 192.9 F 8.2
12:00 207.2 F 11.6
14:00 217.9 F 5.5
16:00 226.7 F 9.6
18:00 234.8 F 11.1
20:00 237.1 F 10.9
22:00 234.9 F 8.7
24:00 247.6 F 14.7
26:00 258.0 F 20.5
28:00 267.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 271.3 F 29.7
2:00 272.1 F 29.7
3:00 272.3 F 29.8
4:00 272.4 F 30.0

Min 270.9 F 29.1
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.1 F 0.0
5:00 229.7 F 0.0
10:00 221.7 F 0.0
15:00 230.0 F 0.0
20:00 224.6 F 0.0
25:00 225.4 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 29:23
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:08:05
CYCLE COMPLETE

09/09/2024 12:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____