
Midmark M11 - v1.0.0
Total Cycles: 1927

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2022 07:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.2 F 0.0
2:00 75.9 F 0.0
4:00 80.2 F 0.0
6:00 102.9 F 1.0
8:00 140.9 F 2.5
10:00 175.0 F 4.6
12:00 195.8 F 7.6
14:00 211.2 F 11.2
16:00 225.6 F 4.1
18:00 226.1 F 7.9
20:00 236.4 F 12.0
22:00 236.3 F 8.4
24:00 244.5 F 11.5
26:00 245.1 F 12.3
28:00 256.9 F 18.4
30:00 267.2 F 24.9
32:00 272.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.2 F 28.2
2:00 272.7 F 28.5
3:00 272.4 F 28.7
4:00 271.7 F 28.1

Min 271.6 F 28.0
Max 272.8 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.0 F 0.0
5:00 183.9 F 0.0
10:00 183.2 F 0.0
15:00 185.9 F 0.0
20:00 186.8 F 0.0
25:00 215.2 F 0.0
30:00 234.9 F 0.0

Drying Complete

FILLING: 0:44
HEATING: 32:13
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:10:02
CYCLE COMPLETE

05/11/2022 08:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____