
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002873
16:35 16/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:28
STERILIZING 3:29
136.4 C 222kPa 3:29
136.4 C 225kPa 3:59
136.1 C 219kPa 4:29
136.4 C 223kPa 4:59
136.4 C 224kPa 5:29
136.4 C 222kPa 5:59
136.4 C 222kPa 6:29
136.4 C 220kPa 6:59
136.4 C 220kPa 6:59
VENTING 6:59
AIR DRYING 7:56
CYCLE COMPLETE 9:49

Digital Signature #
5F0D7123A419710B
