

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3118

Name: RITTER2  
Sterilizer ID: V2313228  
11/13/2023 02:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 149.7 F 0.0  
2:00 152.6 F 2.1  
4:00 169.6 F 5.7  
6:00 191.3 F 9.8  
8:00 207.1 F 7.4  
10:00 217.8 F 8.3  
12:00 230.9 F 13.3  
14:00 236.7 F 10.6  
16:00 235.4 F 9.0  
18:00 249.8 F 15.7  
20:00 261.5 F 22.3  
22:00 272.0 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 29.5  
1:00 272.6 F 29.7  
2:00 272.8 F 29.8  
3:00 272.3 F 30.0  
4:00 271.7 F 29.7

Min 271.6 F 29.3  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 193.4 F 0.0  
10:00 190.0 F 0.0  
15:00 192.2 F 0.0  
20:00 192.0 F 0.0  
25:00 189.8 F 0.0  
30:00 188.6 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 23:06  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:01:48  
CYCLE COMPLETE

11/13/2023 03:37 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_