

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000131  
13:46 08/09/2020  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CYCLE ABORTED 0:03  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 0:03  
-----

Digital Signature #  
9F2A3498C36EB61F  
-----