
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006121
13:34 15/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:32
STERILIZING 3:27
136.4 C 220kPa 3:27
136.2 C 222kPa 3:57
136.3 C 224kPa 4:27
136.3 C 222kPa 4:57
136.4 C 223kPa 5:27
136.4 C 224kPa 5:57
136.4 C 223kPa 6:27
136.3 C 223kPa 6:57
136.3 C 223kPa 6:57
VENTING 6:58
AIR DRYING 7:48
CYCLE COMPLETE 8:14

Digital Signature #
E83D07C87ADD6052
