

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1346

Name: RITTER 1  
Sterilizer ID: V2178198  
11/09/2021 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.8 F 0.0  
2:00 201.7 F 3.2  
4:00 206.7 F 7.9  
6:00 216.7 F 12.6  
8:00 229.6 F 8.4  
10:00 241.5 F 11.6  
12:00 241.5 F 11.6  
14:00 239.7 F 10.2  
16:00 253.3 F 17.0  
18:00 265.2 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.3 F 28.7  
2:00 271.8 F 28.6  
3:00 272.3 F 28.6  
4:00 271.7 F 28.5  
  
Min 271.5 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.3 F 0.0  
5:00 225.7 F 0.0  
10:00 230.9 F 0.0  
15:00 230.6 F 0.0  
20:00 229.6 F 0.0  
25:00 230.6 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 19:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:01  
CYCLE COMPLETE

11/09/2021 01:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_