
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 004805
15:31 05/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:31
STERILIZING 3:28
136.0 C 220kPa 3:28
136.4 C 225kPa 3:58
136.4 C 224kPa 4:28
136.4 C 224kPa 4:58
136.3 C 224kPa 5:28
136.4 C 225kPa 5:58
136.2 C 221kPa 6:28
136.4 C 225kPa 6:58
136.4 C 225kPa 6:58
VENTING 6:58
AIR DRYING 7:52
CYCLE COMPLETE 13:30

Digital Signature #
7241E1BC36AEC28E
