
Midmark M11 - v1.0.0
Total Cycles: 4638

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2024 03:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.0 F 0.0
2:00 129.7 F 1.5
4:00 142.7 F 3.9
6:00 174.2 F 7.0
8:00 201.9 F 10.9
10:00 213.0 F 4.2
12:00 223.4 F 6.7
14:00 233.1 F 10.6
16:00 234.4 F 8.2
18:00 241.0 F 11.1
20:00 235.4 F 8.5
22:00 247.6 F 14.3
24:00 258.1 F 19.9
26:00 267.6 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.9 F 29.3
2:00 271.6 F 28.6
3:00 272.2 F 29.2
4:00 272.2 F 29.1

Min 271.2 F 28.5
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 212.7 F 0.0
10:00 191.6 F 0.0
15:00 189.7 F 0.0
20:00 188.3 F 0.0
25:00 187.7 F 0.0
30:00 187.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:33
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:06:22
CYCLE COMPLETE

06/13/2024 04:28 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____