
Midmark M11 - v1.0.0
Total Cycles: 4148

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.2 F 0.0
2:00 105.0 F 0.8
4:00 149.4 F 2.9
6:00 170.0 F 6.0
8:00 181.0 F 9.5
10:00 195.8 F 13.2
12:00 209.7 F 4.8
14:00 220.5 F 8.6
16:00 231.6 F 12.7
18:00 232.5 F 8.8
20:00 242.5 F 11.7
22:00 243.6 F 11.9
24:00 254.7 F 17.7
26:00 265.0 F 23.9
28:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 271.8 F 29.0
2:00 272.7 F 29.4
3:00 273.0 F 29.3
4:00 272.5 F 29.1

Min 271.0 F 28.3
Max 273.2 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.4 F 0.0
5:00 213.6 F 0.0
10:00 221.8 F 0.0
15:00 217.6 F 0.0
20:00 221.1 F 0.0
25:00 226.6 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 28:11
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:06:53
CYCLE COMPLETE

12/21/2023 02:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____