
Midmark M11 - v1.0.0
Total Cycles: 3499

Name: RITTER 1
Sterilizer ID: V2178198
07/06/2023 03:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.3 F 0.0
2:00 173.1 F 1.4
4:00 193.6 F 5.3
6:00 220.4 F 10.8
8:00 226.5 F 5.9
10:00 237.3 F 11.6
12:00 236.0 F 9.2
14:00 239.3 F 9.2
16:00 249.1 F 14.8
18:00 261.6 F 21.8
20:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.0
2:00 271.4 F 28.3
3:00 272.5 F 29.2
4:00 272.0 F 28.6

Min 271.2 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 217.2 F 0.0
10:00 234.9 F 0.0
15:00 232.5 F 0.0
20:00 223.0 F 0.0
25:00 208.7 F 0.0
30:00 203.5 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:01
CYCLE COMPLETE

07/06/2023 04:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____