
Midmark M11 - v1.0.0
Total Cycles: 3966

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2023 11:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.9 F 0.0
2:00 144.3 F 2.1
4:00 159.7 F 5.4
6:00 193.4 F 9.2
8:00 210.5 F 10.1
10:00 216.9 F 6.8
12:00 228.2 F 11.0
14:00 228.2 F 6.7
16:00 241.2 F 12.4
18:00 239.4 F 10.5
20:00 251.7 F 16.5
22:00 262.6 F 22.8
24:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.8 F 29.2
2:00 271.4 F 28.9
3:00 272.7 F 29.2
4:00 271.2 F 28.7

Min 271.2 F 28.6
Max 273.1 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 217.6 F 0.0
10:00 221.9 F 0.0
15:00 220.6 F 0.0
20:00 217.2 F 0.0
25:00 215.3 F 0.0
30:00 205.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 24:23
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:03:14
CYCLE COMPLETE

11/01/2023 12:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____