

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2528

Name: RITTER 1  
Sterilizer ID: V2178198  
11/08/2022 11:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.0 F 0.0  
2:00 195.8 F 4.3  
4:00 191.2 F 9.0  
6:00 216.3 F 8.1  
8:00 226.1 F 8.3  
10:00 237.0 F 12.0  
12:00 240.1 F 11.0  
14:00 237.8 F 9.2  
16:00 252.3 F 16.2  
18:00 264.4 F 23.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.3 F 28.5  
2:00 271.5 F 28.3  
3:00 272.2 F 28.4  
4:00 271.5 F 28.2  
  
Min 271.4 F 27.9  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.2 F 0.0  
5:00 233.8 F 0.0  
10:00 227.3 F 0.0  
15:00 224.9 F 0.0  
20:00 221.8 F 0.0  
25:00 222.2 F 0.0  
30:00 221.1 F 0.0  
Drying Complete

FILLING: 0:60  
HEATING: 19:53  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:01  
CYCLE COMPLETE

11/08/2022 12:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_