

-----  
Midmark M11 - v1.0.5  
Total Cycles: 196

Name: RITTER2  
Sterilizer ID: V2313228  
07/12/2021 01:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.0 F 0.0  
2:00 199.8 F 4.0  
4:00 208.7 F 9.5  
6:00 224.6 F 8.4  
8:00 236.1 F 11.1  
10:00 235.6 F 9.2  
12:00 239.1 F 9.9  
14:00 249.6 F 15.3  
16:00 261.5 F 22.0  
18:00 272.1 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.7  
1:00 272.6 F 29.4  
2:00 271.5 F 29.0  
3:00 272.3 F 29.1  
4:00 271.6 F 29.1  
  
Min 271.4 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 230.3 F 0.0  
10:00 214.5 F 0.0  
15:00 204.9 F 0.0  
20:00 192.5 F 0.0  
25:00 176.5 F 0.0  
30:00 188.1 F 0.0  
Drying Complete

FILLING: 2:41  
HEATING: 18:22  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

07/12/2021 02:51 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01