

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000267  
13:09 22/07/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:15  
STERILIZING 4:10  
136.0 C 220kPa 4:10  
136.4 C 220kPa 4:41  
136.4 C 223kPa 5:11  
136.3 C 222kPa 5:41  
136.4 C 225kPa 6:11  
136.4 C 222kPa 6:41  
136.0 C 220kPa 7:11  
136.4 C 224kPa 7:41  
136.4 C 224kPa 7:41  
VENTING 7:41  
AIR DRYING 8:32  
CYCLE COMPLETE 8:54

-----  
Digital Signature #  
6C5EAD2A2E994EEB  
-----