
Midmark M11 - v1.0.5
Total Cycles: 3789

Name: RITTER2
Sterilizer ID: V2313228
07/16/2024 01:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 114.2 F 0.0
2:00 116.4 F 1.0
4:00 129.5 F 3.4
6:00 155.3 F 6.5
8:00 179.0 F 10.2
10:00 199.4 F 7.8
12:00 215.0 F 7.3
14:00 226.5 F 11.4
16:00 228.6 F 7.9
18:00 240.7 F 13.1
20:00 240.3 F 11.6
22:00 252.1 F 17.4
24:00 262.6 F 23.6
26:00 272.5 F 30.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 30.0
1:00 272.7 F 30.4
2:00 271.7 F 30.4
3:00 272.3 F 30.2
4:00 271.8 F 30.5

Min 271.4 F 29.9
Max 272.9 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 207.7 F 0.0
10:00 214.8 F 0.0
15:00 214.9 F 0.0
20:00 215.2 F 0.0
25:00 215.5 F 0.0
30:00 222.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 27:13
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:06:05
CYCLE COMPLETE

07/16/2024 02:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____