

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1153

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2022 01:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
2:00 182.8 F 4.0  
4:00 200.9 F 9.2  
6:00 221.1 F 7.7  
8:00 229.1 F 10.4  
10:00 232.7 F 8.0  
12:00 241.0 F 11.1  
14:00 246.4 F 14.0  
16:00 259.3 F 21.1  
18:00 270.6 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.5 F 29.7  
2:00 271.5 F 29.0  
3:00 272.5 F 29.9  
4:00 272.0 F 29.3  
  
Min 271.4 F 28.9  
Max 272.9 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.3 F 0.0  
5:00 220.7 F 0.0  
10:00 225.5 F 0.0  
15:00 224.0 F 0.0  
20:00 223.1 F 0.0  
25:00 220.1 F 0.0  
30:00 218.9 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 18:47  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:53  
CYCLE COMPLETE

06/01/2022 02:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_