
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001244
15:59 01/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:47
STERILIZING 3:55
136.4 C 220kPa 3:55
136.3 C 224kPa 4:25
136.4 C 223kPa 4:55
136.2 C 224kPa 5:25
136.3 C 224kPa 5:55
136.3 C 224kPa 6:25
136.4 C 223kPa 6:55
136.2 C 223kPa 7:25
136.2 C 223kPa 7:25
VENTING 7:25
AIR DRYING 8:16
CYCLE COMPLETE 20:02

Digital Signature #
37ED8D0D26D5FC0E
