
Midmark M11 - v1.0.0
Total Cycles: 1074

Name: RITTER 1
Sterilizer ID: V2178198
08/05/2021 09:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.4 F 0.0
2:00 156.6 F 2.0
4:00 173.1 F 5.5
6:00 195.3 F 10.0
8:00 215.7 F 4.9
10:00 230.2 F 8.8
12:00 238.8 F 11.8
14:00 239.4 F 10.8
16:00 236.3 F 8.6
18:00 250.5 F 15.4
20:00 262.2 F 22.1
22:00 272.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.5 F 28.7
2:00 272.4 F 28.9
3:00 272.1 F 28.4
4:00 272.7 F 29.1

Min 271.7 F 28.2
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 204.9 F 0.0
10:00 193.6 F 0.0
15:00 194.4 F 0.0
20:00 190.7 F 0.0
25:00 195.6 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

08/05/2021 12:25 PM
ACCEPTED: X
REJECTED:
Operator: U02