
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000979
9:44 10/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:15
STERILIZING 3:15
136.0 C 220kPa 3:15
136.3 C 232kPa 3:45
136.4 C 233kPa 4:15
136.4 C 233kPa 4:45
136.4 C 233kPa 5:15
136.4 C 232kPa 5:45
136.4 C 232kPa 6:15
136.4 C 231kPa 6:45
136.4 C 231kPa 6:45
VENTING 6:45
AIR DRYING 7:35
CYCLE COMPLETE 7:37

Digital Signature #
49D28FE37B09BE66
