
Midmark M11 - v1.0.5
Total Cycles: 1266

Name: RITTER2
Sterilizer ID: V2313228
07/07/2022 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 163.0 F 1.8
4:00 185.7 F 5.6
6:00 191.8 F 11.1
8:00 215.9 F 5.5
10:00 227.1 F 10.8
12:00 230.1 F 7.3
14:00 242.4 F 13.3
16:00 240.9 F 11.8
18:00 253.2 F 18.0
20:00 264.1 F 24.4
22:00 272.1 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 272.0 F 30.3
2:00 272.6 F 30.0
3:00 271.8 F 29.5
4:00 271.3 F 29.5

Min 271.2 F 29.3
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.3 F 0.0
5:00 235.9 F 0.0
10:00 221.9 F 0.0
15:00 221.1 F 0.0
20:00 219.0 F 0.0
25:00 218.8 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 22:12
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:03
CYCLE COMPLETE

07/07/2022 11:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____