
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002562
10:56 28/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:58
PRESSURIZING 3:01
STERILIZING 3:52
136.0 C 220kPa 3:52
136.4 C 225kPa 4:22
136.4 C 225kPa 4:52
136.2 C 222kPa 5:22
136.3 C 224kPa 5:52
136.4 C 222kPa 6:22
136.3 C 222kPa 6:52
136.1 C 221kPa 7:22
136.1 C 221kPa 7:22
VENTING 7:22
AIR DRYING 8:17
CYCLE COMPLETE 8:22

Digital Signature #
8F9326F85B8F7F6E
