

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000789  
7:12 11/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:20  
STERILIZING 4:26  
136.0 C 220kPa 4:26  
136.3 C 220kPa 4:56  
136.1 C 221kPa 5:26  
136.2 C 221kPa 5:56  
136.3 C 222kPa 6:26  
136.4 C 223kPa 6:56  
136.4 C 222kPa 7:26  
136.4 C 221kPa 7:56  
136.4 C 221kPa 7:56  
VENTING 7:56  
AIR DRYING 8:50  
CYCLE COMPLETE 8:53

-----  
Digital Signature #  
4F0D1C4E1C1110DC  
-----