

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000951  
18:38 23/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:08  
PRESSURIZING 3:15  
STERILIZING 4:55  
136.4 C 232kPa 4:55  
136.4 C 233kPa 5:25  
136.2 C 230kPa 5:55  
136.4 C 232kPa 6:25  
136.4 C 233kPa 6:55  
136.3 C 233kPa 7:25  
136.4 C 233kPa 7:55  
136.3 C 233kPa 8:25  
136.3 C 233kPa 8:25  
VENTING 8:25  
CYCLE ABORTED 8:44  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 8:44  
-----

Digital Signature #  
2BA4806431BB21F6  
-----