
Midmark M11 - v1.0.0
Total Cycles: 1052

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2021 10:24 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.4 F 0.0
2:00 172.8 F 2.1
4:00 189.2 F 5.3
6:00 223.4 F 9.9
8:00 221.3 F 6.8
10:00 233.3 F 8.9
12:00 243.1 F 12.2
14:00 240.2 F 11.0
16:00 237.1 F 9.1
18:00 250.4 F 15.5
20:00 261.4 F 21.8
22:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.1
2:00 271.6 F 28.3
3:00 272.4 F 29.1
4:00 272.4 F 28.9

Min 271.4 F 28.3
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 212.9 F 0.0
10:00 215.8 F 0.0
15:00 215.0 F 0.0
20:00 212.9 F 0.0
25:00 216.7 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 1:22
HEATING: 22:37
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:22
CYCLE COMPLETE

07/26/2021 11:25 AM
ACCEPTED: X
REJECTED:
Operator: U01