

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2703

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2023 09:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.2 F 0.0  
2:00 143.3 F 0.0  
4:00 159.3 F 0.2  
6:00 185.6 F 4.1  
8:00 204.8 F 8.5  
10:00 217.5 F 13.2  
12:00 225.9 F 8.1  
14:00 237.5 F 13.0  
16:00 238.3 F 10.6  
18:00 235.8 F 8.8  
20:00 250.2 F 15.5  
22:00 261.4 F 22.0  
24:00 271.6 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.6 F 29.2  
2:00 271.6 F 28.8  
3:00 272.7 F 29.3  
4:00 271.6 F 28.6

Min 271.5 F 28.6  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
5:00 194.5 F 0.0  
10:00 194.0 F 0.0  
15:00 192.5 F 0.0  
20:00 189.0 F 0.0  
25:00 186.4 F 0.0  
30:00 185.9 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 24:37  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:02:47  
CYCLE COMPLETE

07/26/2023 10:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_