

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002003  
11:28 10/09/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:52  
PRESSURIZING 2:00  
STERILIZING 3:01  
136.4 C 226kPa 3:01  
136.3 C 225kPa 3:31  
136.1 C 223kPa 4:01  
136.2 C 225kPa 4:31  
136.3 C 225kPa 5:01  
136.3 C 226kPa 5:31  
136.4 C 227kPa 6:01  
136.1 C 223kPa 6:31  
136.1 C 223kPa 6:31  
VENTING 6:31  
AIR DRYING 7:22  
CYCLE COMPLETE 8:00  
-----

Digital Signature #  
5C9C768615AC37EE  
-----