
Midmark M11 - v1.0.0
Total Cycles: 1983

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2022 03:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.5 F 0.0
2:00 166.3 F 3.5
4:00 186.9 F 8.0
6:00 222.6 F 11.3
8:00 226.2 F 7.7
10:00 237.4 F 12.8
12:00 236.3 F 9.6
14:00 239.4 F 8.7
16:00 248.4 F 14.4
18:00 260.7 F 21.3
20:00 271.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 29.1
2:00 271.7 F 28.3
3:00 272.2 F 29.1
4:00 272.4 F 28.8

Min 271.4 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 214.6 F 0.0
10:00 223.7 F 0.0
15:00 228.8 F 0.0
20:00 231.4 F 0.0
25:00 231.5 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 20:34
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

05/31/2022 04:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____