
Midmark M11 - v1.0.5
Total Cycles: 1501

Name: RITTER2
Sterilizer ID: V2313228
09/19/2022 11:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.5 F 0.0
2:00 163.4 F 2.8
4:00 172.2 F 6.5
6:00 193.4 F 11.4
8:00 217.4 F 5.9
10:00 229.7 F 11.1
12:00 230.9 F 8.1
14:00 240.7 F 11.7
16:00 244.3 F 13.1
18:00 256.9 F 19.8
20:00 268.1 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.4
1:00 272.7 F 30.0
2:00 271.7 F 29.3
3:00 272.1 F 29.9
4:00 272.5 F 29.9

Min 271.2 F 29.2
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 212.7 F 0.0
10:00 230.2 F 0.0
15:00 226.7 F 0.0
20:00 229.2 F 0.0
25:00 229.0 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

09/19/2022 12:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____