
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002639
12:53 31/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:20
PRESSURIZING 2:27
STERILIZING 3:30
136.0 C 220kPa 3:30
136.3 C 225kPa 4:00
136.4 C 222kPa 4:30
136.4 C 223kPa 5:00
136.2 C 223kPa 5:30
136.4 C 226kPa 6:00
136.4 C 223kPa 6:30
136.4 C 222kPa 7:00
136.4 C 222kPa 7:00
VENTING 7:00
AIR DRYING 7:51
CYCLE COMPLETE 12:06

Digital Signature #
00FC88AA3FBDA80E
