
Midmark M11 - v1.0.0
Total Cycles: 676

Name: MIDMARK
Sterilizer ID: V2178198
02/18/2021 10:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 118.8 F 0.0
2:00 145.0 F 0.1
4:00 134.7 F 1.9
6:00 170.0 F 5.3
8:00 195.5 F 10.1
10:00 226.6 F 5.0
12:00 223.5 F 9.0
14:00 236.7 F 10.0
16:00 242.9 F 11.5
18:00 239.1 F 9.6
20:00 253.1 F 16.5
22:00 265.0 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.9 F 28.9
2:00 271.3 F 28.3
3:00 272.5 F 28.9
4:00 271.3 F 28.1

Min 271.2 F 28.1
Max 273.1 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 235.5 F 0.0
10:00 233.5 F 0.0
15:00 226.2 F 0.0
20:00 224.3 F 0.0
25:00 227.9 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:50
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:01:49
CYCLE COMPLETE

02/18/2021 11:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____