
Midmark M11 - v1.0.5
Total Cycles: 2626

Name: RITTER2
Sterilizer ID: V2313228
07/10/2023 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 188.0 F 3.1
4:00 204.9 F 8.0
6:00 218.8 F 12.3
8:00 230.3 F 9.3
10:00 235.6 F 9.6
12:00 244.1 F 13.3
14:00 243.7 F 12.5
16:00 256.6 F 19.4
18:00 268.0 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.4 F 29.3
2:00 272.2 F 29.7
3:00 271.9 F 28.9
4:00 272.7 F 29.7

Min 271.5 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 231.0 F 0.0
10:00 226.8 F 0.0
15:00 221.1 F 0.0
20:00 216.5 F 0.0
25:00 215.8 F 0.0
30:00 221.5 F 0.0
Drying Complete

FILLING: 2:10
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

07/10/2023 11:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____