

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001727  
15:53 27/03/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:28  
STERILIZING 3:31  
136.0 C 220kPa 3:31  
136.4 C 226kPa 4:01  
136.2 C 221kPa 4:31  
136.4 C 221kPa 5:01  
136.3 C 225kPa 5:31  
136.2 C 224kPa 6:01  
136.4 C 225kPa 6:31  
136.4 C 226kPa 7:01  
136.4 C 226kPa 7:01  
VENTING 7:01  
AIR DRYING 7:50  
CYCLE COMPLETE 11:18  
-----

Digital Signature #  
D7F8ED00F777C0D5  
-----