
Midmark M11 - v1.0.0
Total Cycles: 800

Name: RITTER 1
Sterilizer ID: V2178198
04/22/2021 10:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.1 F 0.0
2:00 194.4 F 3.0
4:00 209.5 F 7.9
6:00 216.8 F 13.1
8:00 225.2 F 8.3
10:00 237.5 F 13.5
12:00 239.7 F 11.3
14:00 239.4 F 10.0
16:00 252.5 F 16.6
18:00 263.9 F 23.3
20:00 271.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.3 F 28.9
2:00 272.4 F 29.2
3:00 272.7 F 29.4
4:00 271.8 F 29.0

Min 271.8 F 28.8
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.4 F 0.0
5:00 247.8 F 0.0
10:00 227.1 F 0.0
15:00 231.5 F 0.0
20:00 227.9 F 0.0
25:00 216.8 F 0.0
30:00 215.3 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:00:13
CYCLE COMPLETE

04/22/2021 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____