
Midmark M11 - v1.0.0
Total Cycles: 986

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2021 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 184.3 F 3.4
4:00 208.4 F 8.7
6:00 220.3 F 8.7
8:00 234.2 F 9.9
10:00 234.3 F 8.1
12:00 244.7 F 12.4
14:00 245.8 F 12.9
16:00 259.2 F 20.3
18:00 271.3 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.0 F 28.3
2:00 272.5 F 28.9
3:00 271.3 F 28.0
4:00 272.5 F 28.7

Min 271.3 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.4 F 0.0
5:00 204.2 F 0.0
10:00 208.9 F 0.0
15:00 207.7 F 0.0
20:00 205.7 F 0.0
25:00 221.1 F 0.0
30:00 193.1 F 0.0
Drying Complete

FILLING: 2:33
HEATING: 18:36
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:14
CYCLE COMPLETE

07/08/2021 02:22 PM
ACCEPTED: X
REJECTED:
Operator: U02