

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2399

Name: RITTER2  
Sterilizer ID: V2313228  
05/23/2023 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.7 F 0.0  
2:00 174.1 F 3.4  
4:00 186.9 F 8.0  
6:00 208.8 F 13.5  
8:00 224.3 F 8.1  
10:00 236.1 F 13.0  
12:00 238.6 F 11.3  
14:00 237.8 F 9.9  
16:00 251.4 F 16.6  
18:00 263.2 F 23.5  
20:00 272.1 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 272.6 F 29.7  
2:00 271.7 F 29.4  
3:00 272.6 F 29.6  
4:00 271.6 F 29.2

Min 271.6 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.8 F 0.0  
5:00 219.5 F 0.0  
10:00 239.1 F 0.0  
15:00 235.3 F 0.0  
20:00 225.5 F 0.0  
25:00 222.5 F 0.0  
30:00 225.9 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 20:09  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:08  
CYCLE COMPLETE

05/23/2023 12:00 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_