
Midmark M11 - v1.0.0
Total Cycles: 3471

Name: RITTER 1
Sterilizer ID: V2178198
06/30/2023 10:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.9 F 0.0
2:00 196.8 F 2.5
4:00 193.9 F 6.5
6:00 213.7 F 11.2
8:00 221.5 F 5.8
10:00 232.8 F 10.6
12:00 233.1 F 7.7
14:00 245.0 F 13.3
16:00 243.8 F 12.0
18:00 256.0 F 18.3
20:00 266.7 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.5 F 28.8
2:00 271.8 F 28.1
3:00 271.8 F 28.5
4:00 272.5 F 28.8

Min 271.4 F 27.9
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 213.7 F 0.0
10:00 222.1 F 0.0
15:00 221.1 F 0.0
20:00 217.9 F 0.0
25:00 217.5 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 21:37
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

06/30/2023 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____