
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001295
12:48 13/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:43
STERILIZING 3:54
136.0 C 220kPa 3:54
136.2 C 223kPa 4:24
136.2 C 224kPa 4:54
136.1 C 222kPa 5:24
136.4 C 225kPa 5:54
136.4 C 222kPa 6:24
136.3 C 224kPa 6:54
136.4 C 224kPa 7:24
136.4 C 224kPa 7:24
VENTING 7:24
AIR DRYING 8:17
CYCLE COMPLETE 23:17

Digital Signature #
91CD3092EB21BF52
