
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001610
11:24 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:11
STERILIZING 4:18
136.0 C 223kPa 4:18
136.3 C 225kPa 4:48
136.4 C 226kPa 5:18
136.4 C 223kPa 5:48
136.3 C 225kPa 6:18
136.3 C 225kPa 6:48
136.4 C 226kPa 7:18
136.4 C 226kPa 7:48
136.4 C 226kPa 7:48
VENTING 7:48
AIR DRYING 8:43
CYCLE COMPLETE 8:49

Digital Signature #
617A4A4AF69C4B09
