

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001935  
12:59 08/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:52  
PRESSURIZING 1:58  
STERILIZING 2:58  
136.4 C 224kPa 2:58  
136.2 C 225kPa 3:29  
136.3 C 225kPa 3:59  
136.0 C 222kPa 4:29  
136.4 C 225kPa 4:59  
136.4 C 224kPa 5:29  
136.4 C 226kPa 5:59  
136.4 C 224kPa 6:29  
136.4 C 224kPa 6:29  
VENTING 6:29  
AIR DRYING 7:20  
CYCLE COMPLETE 37:20  
-----

Digital Signature #  
55A9726CBADD839F  
-----