
Midmark M11 - v1.0.5
Total Cycles: 4209

Name: RITTER2
Sterilizer ID: V2313228
12/12/2024 05:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.8 F 0.0
2:00 110.5 F 0.0
4:00 129.8 F 0.1
6:00 157.6 F 3.7
8:00 185.4 F 7.9
10:00 206.0 F 12.6
12:00 216.6 F 6.3
14:00 229.9 F 10.9
16:00 230.3 F 7.3
18:00 243.6 F 13.2
20:00 243.2 F 12.3
22:00 256.0 F 19.1
24:00 267.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.0
1:00 272.7 F 29.9
2:00 271.8 F 29.0
3:00 272.8 F 29.6
4:00 271.9 F 29.2

Min 271.6 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.7 F 0.0
5:00 210.8 F 0.0
10:00 223.4 F 0.0
15:00 218.3 F 0.0
20:00 216.8 F 0.0
25:00 217.9 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:11
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:03:17
CYCLE COMPLETE

12/12/2024 06:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____