
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001902
14:47 28/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:31
PRESSURIZING 3:31
STERILIZING 4:29
136.4 C 222kPa 4:29
136.4 C 224kPa 4:59
136.5 C 225kPa 5:29
136.4 C 224kPa 5:59
136.4 C 224kPa 6:29
136.4 C 225kPa 6:59
136.4 C 225kPa 7:29
136.4 C 225kPa 7:59
136.4 C 225kPa 7:59
VENTING 7:59
AIR DRYING 8:54
CYCLE COMPLETE 8:58

Digital Signature #
BD57122C60338306
