
Midmark M11 - v1.0.0
Total Cycles: 4812

Name: RITTER 1
Sterilizer ID: V2178198
08/13/2024 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.4 F 0.0
2:00 165.7 F 4.0
4:00 191.3 F 9.1
6:00 207.0 F 4.8
8:00 226.9 F 8.6
10:00 236.4 F 10.3
12:00 240.5 F 11.3
14:00 238.2 F 9.4
16:00 252.8 F 16.6
18:00 265.4 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.4 F 28.4
2:00 272.2 F 28.6
3:00 272.0 F 28.1
4:00 272.6 F 28.8

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 202.5 F 0.0
10:00 210.9 F 0.0
15:00 222.7 F 0.0
20:00 227.0 F 0.0
25:00 234.4 F 0.0
30:00 228.6 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

08/13/2024 04:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____