

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2567

Name: RITTER 1  
Sterilizer ID: V2178198  
11/16/2022 12:37 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.5 F 0.0  
2:00 203.9 F 4.5  
4:00 211.7 F 9.4  
6:00 226.8 F 6.7  
8:00 230.9 F 9.7  
10:00 233.3 F 8.6  
12:00 243.0 F 12.8  
14:00 242.1 F 11.6  
16:00 255.4 F 18.4  
18:00 267.2 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 29.2  
2:00 271.6 F 28.4  
3:00 272.3 F 29.3  
4:00 272.3 F 29.0  
  
Min 271.5 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 233.0 F 0.0  
10:00 236.4 F 0.0  
15:00 229.2 F 0.0  
20:00 227.0 F 0.0  
25:00 227.1 F 0.0  
30:00 225.0 F 0.0  
  
Drying Complete

FILLING: 1:33  
HEATING: 19:26  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

11/16/2022 01:35 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_