
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 000292
8:11 05/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:14
STERILIZING 4:11
136.0 C 220kPa 4:11
136.4 C 224kPa 4:41
136.4 C 224kPa 5:11
136.4 C 222kPa 5:41
136.2 C 222kPa 6:11
136.1 C 222kPa 6:41
136.4 C 225kPa 7:11
136.4 C 224kPa 7:41
136.4 C 224kPa 7:41
VENTING 7:41
AIR DRYING 8:30
CYCLE COMPLETE 9:01

Digital Signature #
59E0CF5505A63B1A
