
Midmark M11 - v1.0.0
Total Cycles: 5102

Name: RITTER 1
Sterilizer ID: V2178198
11/18/2024 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.1 F 0.0
2:00 154.7 F 0.0
4:00 160.2 F 0.3
6:00 181.9 F 1.8
8:00 201.0 F 5.3
10:00 215.8 F 9.0
12:00 225.1 F 13.3
14:00 230.4 F 8.6
16:00 240.8 F 13.4
18:00 239.9 F 10.9
20:00 236.6 F 9.1
22:00 250.4 F 15.5
24:00 261.4 F 21.7
26:00 271.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 29.1
2:00 272.2 F 28.8
3:00 271.4 F 28.3
4:00 271.5 F 28.7

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 198.4 F 0.0
10:00 207.6 F 0.0
15:00 223.9 F 0.0
20:00 220.9 F 0.0
25:00 224.0 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 26:46
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:05:21
CYCLE COMPLETE

11/18/2024 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____