
Midmark M11 - v1.0.0
Total Cycles: 1453

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2021 09:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.3 F 0.0
2:00 159.9 F 1.4
4:00 174.9 F 3.9
6:00 189.5 F 7.5
8:00 207.8 F 12.3
10:00 219.5 F 5.9
12:00 230.9 F 10.8
14:00 231.4 F 6.8
16:00 243.3 F 12.8
18:00 241.5 F 11.0
20:00 253.8 F 17.3
22:00 264.9 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 28.9
2:00 271.4 F 28.3
3:00 272.6 F 29.0
4:00 271.5 F 28.2

Min 271.4 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.6 F 0.0
5:00 215.5 F 0.0
10:00 224.9 F 0.0
15:00 223.7 F 0.0
20:00 223.9 F 0.0
25:00 229.7 F 0.0
30:00 232.8 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:56
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:02:15
CYCLE COMPLETE

12/14/2021 10:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____