

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002310  
12:01 20/11/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:57  
PRESSURIZING 2:03  
STERILIZING 3:08  
136.0 C 221kPa 3:08  
136.2 C 224kPa 3:38  
136.4 C 226kPa 4:08  
136.4 C 226kPa 4:38  
136.4 C 226kPa 5:08  
136.1 C 224kPa 5:38  
136.4 C 223kPa 6:08  
136.3 C 223kPa 6:38  
136.3 C 223kPa 6:38  
VENTING 6:38  
AIR DRYING 7:28  
CYCLE COMPLETE 37:28  
-----

Digital Signature #  
EE31196A028D7F25  
-----